Gilson Test Report

IDENTIFICATION

Model:	Pipetman L P200L
Useful volume range:	20 µL to 200 µL
Serial N°:	TM73678
Tips:	Diamond D200

TEST CONDITIONS

Balance n°:	91
Sensitivity:	10^{-5} g
Correction factor Z (µL/mg):	1.0032
Correction for evaporation:	0 µL
Basis of adjustment:	Ex
Air temperature:	21.5°C
Barometric pressure:	101 kPa
Relative humidity:	60 %

CONTROL

Date: 14 NOV 2022 **Inspected by:** DD

VOLUMETRIC DATA

Gravimetric results are converted to volumetric data using the corresponding correction factor Z and taking into account the evaporation if applicable.

Mean:	20.15 µL
Systematic error:	0.15 µL
es %:	0.75 %
Random error:	0.03 µL
CV%:	0.15 %

TEST RESULTS

Selected volume:	20 µL – PASS
Nominal volume:	200 µL – PASS

WARRANTY PERIOD

3 years

Gilson Warranty

Gilson warrants this pipette against defects in material under normal use and service for the period specified in the warranty document, from the date of purchase. This warranty shall not apply to pipettes which are subject to abnormal use and/or improper or inadequate maintenance (contrary to the recommendations given in the u ser's guide), including, but not limited to pipettes which have been subjected to physical damage, improper handling, or spillage or exposure to any corrosive environment. This warranty shall also be void in the event pipettes are altered or modified by any party other than Gilson or its designates. Gilson's sole liability under this warranty shall be limited to, at Gilson's sole option, repair or replacement of any defective components of pipettes or refund of the purchase price paid for such pipettes.

The foregoing warranty is exclusive and Gilson hereby disclaims all other warranties, wether express or implied, including any warranties of merchantability and any warranties of fitness for a particular purpose, under no circumstances shall Gilson be liable for any consequential, punitive, indirect or incidental damages arising out of any breach of any express or implied warranty.

GILSON®

LT803115/H - © June 2016