

Gilson Test Report

Gilson Warranty

Gilson warrants this pipette against defects in material under normal use and service for the period specified in the warranty document, from the date of purchase. This warranty shall not apply to pipettes which are subject to abnormal use and/or improper or inadequate maintenance (contrary to the recommendations given in the user's guide), including, but not limited to pipettes which have been subjected to physical damage, improper handling, or spillage or exposure to any corrosive environment. This warranty shall also be void in the event pipettes are altered or modified by any party other than Gilson or its assignates. Gilson's sole liability under this warranty shall be limited to, or Gilson's sole option, repair or replacement of any defective components of pipettes or refund of the purchase price paid for such pipettes.

The foregoing warranty is exclusive and Gilson hereby disclaims all other warranties, whether express or implied, including any warranties of merchantability and any warranties of fitness for a particular purpose, under no circumstances shall Gilson be liable for any consequential, punitive, indirect or incidental damages arising out of any breach of any express or implied warranty.



LT803115/H - © June 2016

<p>VOLUMETRIC DATA</p> <p><i>Gravimetric results are converted to volumetric data using the corresponding correction factor Z and taking into account the evaporation if applicable.</i></p> <p>Mean: 1.982 μL Systematic error: -0.018 μL es %: -0.90 % Random error: 0.009 μL CV%: 0.45 %</p>	
<p>TEST RESULTS</p> <p>Selected volume: 2 μL -- PASS Nominal volume: 20 μL -- PASS</p>	
<p>WARRANTY PERIOD 3 years</p>	

<p>IDENTIFICATION</p> <p>Model: Pipetman L P20L Useful volume range: 2 μL to 20 μL Serial N°: TK71534 Tips: Diamond D200</p>	<p>TEST CONDITIONS</p> <p>Balance n°: 88 Sensitivity: 10^{-6} g Correction factor Z ($\mu\text{L}/\text{mg}$): 1.0032 Correction for evaporation: 0.018 μL Basis of adjustment: Ex Air temperature: 21.5°C Barometric pressure: 101 kPa Relative humidity: 60 %</p>
<p>CONTROL</p> <p>Date: 5 SEP 2022</p>	<p>inspected by: LM</p>