

Gilson Test Report

Gilson Warranty

Gilson warrants this pipette against defects in material under normal use and service for the period specified in the warranty document, from the date of purchase. This warranty shall not apply to pipettes which are subject to abnormal use and/or improper or inadequate maintenance (contrary to the recommendations given in the user's guide), including, but not limited to pipettes which have been subjected to physical damage, improper handling, or spillage or exposure to any corrosive environment. This warranty shall also be void in the event pipettes are altered or modified by any party other than Gilson or its designates. Gilson's sole liability under this warranty shall be limited to, at Gilson's sole option, repair or replacement of any defective components of pipettes or refund of the purchase price paid for such pipettes.

The foregoing warranty is exclusive and Gilson hereby disclaims all other warranties, whether express or implied, including any warranties of merchantability and any warranties of fitness for a particular purpose, under no circumstances shall Gilson be liable for any consequential, punitive, indirect or incidental damages arising out of any breach of any express or implied warranty.



LT803115/H - © June 2016

IDENTIFICATION		VOLUMETRIC DATA	
Model: Pipetman L P20L		Gravimetric results are converted to volumetric data using the corresponding correction factor Z and taking into account the evaporation if applicable.	
Useful volume range: 2 µL to 20 µL		Mean: 2.000 µL	
Serial N°: TK71518		Systematic error: 0.000 µL	
Tips: Diamond D200		es %: 0.00 %	
		Random error: 0.014 µL	
		CV%: 0.70 %	
TEST CONDITIONS		TEST RESULTS	
Balance n°: 88		Selected volume: 2 µL - PASS	
Sensitivity: 10 ⁻⁶ g		Nominal volume: 20 µL - PASS	
Correction factor Z (µL/mg): 1.0032			
Correction for evaporation: 0.018 µL			
Basis of adjustment: Ex			
Air temperature: 21.5°C			
Barometric pressure: 101 kPa			
Relative humidity: 60 %			
CONTROL		WARRANTY PERIOD	
Date: 6 SEP 2022		3 years	
Inspected by: LM			